

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2205

Name: RITTER 1  
Sterilizer ID: V2178198  
08/08/2022 09:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.5 F 0.0  
2:00 131.0 F 0.1  
4:00 149.6 F 2.8  
6:00 173.7 F 6.5  
8:00 203.4 F 10.6  
10:00 212.1 F 4.1  
12:00 225.5 F 8.4  
14:00 237.2 F 13.1  
16:00 236.1 F 9.7  
18:00 239.0 F 8.9  
20:00 246.7 F 13.7  
22:00 258.3 F 20.0  
24:00 269.2 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 28.4  
1:00 271.9 F 28.9  
2:00 272.4 F 29.0  
3:00 271.5 F 28.3  
4:00 272.2 F 29.2

Min 271.3 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 214.9 F 0.0  
10:00 226.7 F 0.0  
15:00 231.0 F 0.0  
20:00 231.9 F 0.0  
25:00 223.5 F 0.0  
30:00 222.9 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 24:59  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 01:02:59  
CYCLE COMPLETE

08/08/2022 10:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_