
Midmark M11 - v1.0.5
Total Cycles: 2486

Name: RITTER2
Sterilizer ID: V2313228
06/08/2023 11:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.0 F 0.0
2:00 169.2 F 3.4
4:00 180.5 F 7.5
6:00 201.6 F 12.7
8:00 219.6 F 7.5
10:00 235.1 F 13.1
12:00 236.8 F 10.7
14:00 235.9 F 9.3
16:00 250.4 F 16.2
18:00 262.5 F 23.2
20:00 272.7 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.7 F 30.1
2:00 271.9 F 29.4
3:00 271.6 F 29.6
4:00 272.8 F 30.0

Min 271.2 F 29.1
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 224.0 F 0.0
10:00 226.9 F 0.0
15:00 228.6 F 0.0
20:00 225.6 F 0.0
25:00 225.7 F 0.0
30:00 224.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

06/08/2023 12:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____