
Midmark M11 - v1.0.5
Total Cycles: 2541

Name: RITTER2
Sterilizer ID: V2313228
06/20/2023 03:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.6 F 0.0
2:00 189.7 F 4.0
4:00 199.3 F 8.9
6:00 218.0 F 8.0
8:00 231.1 F 10.2
10:00 233.3 F 8.1
12:00 241.7 F 11.4
14:00 246.9 F 13.8
16:00 259.8 F 20.9
18:00 271.2 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.5 F 29.2
2:00 271.0 F 28.6
3:00 272.6 F 29.3
4:00 272.2 F 29.0

Min 271.0 F 28.4
Max 273.0 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 218.5 F 0.0
10:00 236.1 F 0.0
15:00 223.0 F 0.0
20:00 220.9 F 0.0
25:00 222.6 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:40
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

06/20/2023 04:09 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____