

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000009  
15:11 13/12/2019  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 0:48  
PRESSURIZING 1:00  
STERILIZING 1:49  
136.0 C 225kPa 1:49  
136.3 C 228kPa 2:19  
136.4 C 226kPa 2:49  
136.3 C 225kPa 3:19  
136.0 C 223kPa 3:49  
136.4 C 223kPa 4:19  
136.2 C 224kPa 4:49  
136.4 C 225kPa 5:19  
136.4 C 225kPa 5:19  
VENTING 5:19  
AIR DRYING 6:06  
CYCLE COMPLETE 6:09

-----  
Digital Signature #  
CDF730070FB05858  
-----