

-----  
Midmark M11 - v1.0.5  
Total Cycles: 378

Name: RITTER2  
Sterilizer ID: V2313228  
09/20/2021 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
2:00 179.7 F 2.7  
4:00 178.4 F 7.0  
6:00 192.0 F 12.4  
8:00 214.2 F 6.2  
10:00 230.0 F 11.7  
12:00 236.9 F 9.5  
14:00 238.1 F 9.3  
16:00 250.0 F 15.2  
18:00 261.9 F 22.1  
20:00 272.3 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 29.2  
1:00 272.2 F 29.7  
2:00 272.6 F 29.5  
3:00 272.2 F 29.3  
4:00 271.8 F 29.2

Min 271.8 F 29.1  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 239.6 F 0.0  
10:00 226.1 F 0.0  
15:00 216.4 F 0.0  
20:00 196.6 F 0.0  
25:00 187.6 F 0.0  
30:00 174.8 F 0.0  
Drying Complete

FILLING: 1:15  
HEATING: 20:59  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 00:59:14  
CYCLE COMPLETE

09/20/2021 10:12 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01