

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000939  
15:02 17/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:45  
PRESSURIZING 2:52  
STERILIZING 3:58  
136.0 C 227kPa 3:58  
136.4 C 232kPa 4:28  
136.4 C 233kPa 4:58  
136.3 C 232kPa 5:28  
136.4 C 232kPa 5:58  
136.3 C 232kPa 6:28  
136.3 C 232kPa 6:58  
136.3 C 232kPa 7:28  
136.3 C 232kPa 7:28  
VENTING 7:28  
AIR DRYING 8:20  
CYCLE COMPLETE 23:20  
-----

Digital Signature #  
5E3533CFE7009573  
-----