
Midmark M11 - v1.0.0
Total Cycles: 988

Name: RITTER 1
Sterilizer ID: V2178198
07/08/2021 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

C232: FILL MODE
WATER LOW
CHAMBER TEMP 205F
CHAMBER PRESS 0.0PSI
** ITEMS NOT STERILE **

07/08/2021 03:19 PM