
Midmark M11 - v1.0.0
Total Cycles: 3606

Name: RITTER 1
Sterilizer ID: V2178198
08/02/2023 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.4 F 0.0
2:00 149.6 F 0.0
4:00 154.5 F 0.7
6:00 188.2 F 4.5
8:00 203.3 F 9.2
10:00 222.0 F 11.0
12:00 228.1 F 7.9
14:00 238.7 F 12.6
16:00 236.9 F 9.6
18:00 238.6 F 9.2
20:00 247.3 F 13.7
22:00 258.8 F 20.0
24:00 269.0 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.7 F 28.9
2:00 271.5 F 28.1
3:00 272.4 F 28.9
4:00 272.2 F 28.5

Min 271.2 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.1 F 0.0
5:00 205.8 F 0.0
10:00 219.9 F 0.0
15:00 219.4 F 0.0
20:00 221.2 F 0.0
25:00 222.7 F 0.0
30:00 212.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 25:10
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:03:17
CYCLE COMPLETE

08/02/2023 11:17 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____