
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 004436
12:17 26/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
PRESSURIZING 3:05
STERILIZING 4:13
136.5 C 221kPa 4:13
136.4 C 226kPa 4:43
136.4 C 225kPa 5:13
136.4 C 225kPa 5:43
136.2 C 223kPa 6:13
136.3 C 222kPa 6:43
136.1 C 219kPa 7:13
136.1 C 220kPa 7:43
136.1 C 220kPa 7:43
VENTING 7:43
AIR DRYING 8:42
CYCLE COMPLETE 8:51

Digital Signature #
7BB051D1F504C110
