

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001393  
8:33 25/03/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:40  
PRESSURIZING 2:44  
STERILIZING 3:46  
136.4 C 223kPa 3:46  
136.0 C 221kPa 4:16  
136.1 C 222kPa 4:46  
136.4 C 224kPa 5:16  
136.2 C 220kPa 5:46  
136.4 C 224kPa 6:16  
136.4 C 224kPa 6:46  
136.1 C 222kPa 7:16  
136.1 C 222kPa 7:16  
VENTING 7:16  
AIR DRYING 8:08  
CYCLE COMPLETE 8:28

-----  
Digital Signature #  
351868795B76EC32  
-----