
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001281
12:35 03/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:29
PRESSURIZING 4:29
STERILIZING 5:46
136.0 C 220kPa 5:46
136.4 C 223kPa 6:16
136.4 C 224kPa 6:46
136.4 C 224kPa 7:16
136.4 C 224kPa 7:46
136.3 C 224kPa 8:16
136.4 C 225kPa 8:46
136.2 C 221kPa 9:16
136.2 C 221kPa 9:16
VENTING 9:16
AIR DRYING 10:14
CYCLE COMPLETE 25:14

Digital Signature #
0E9140D82CF278BE
