
Midmark M11 - v1.0.5
Total Cycles: 847

Name: RITTER2
Sterilizer ID: V2313228
02/28/2022 04:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.4 F 0.0
2:00 160.7 F 3.4
4:00 178.5 F 7.6
6:00 202.0 F 12.8
8:00 219.9 F 7.1
10:00 234.8 F 12.5
12:00 236.6 F 10.1
14:00 236.3 F 8.7
16:00 250.4 F 15.8
18:00 262.5 F 22.9
20:00 272.6 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.7
1:00 272.6 F 29.6
2:00 272.0 F 29.0
3:00 271.2 F 28.9
4:00 272.4 F 29.7

Min 271.0 F 28.6
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 237.9 F 0.0
10:00 210.1 F 0.0
15:00 196.3 F 0.0
20:00 195.8 F 0.0
25:00 221.3 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

03/01/2022 07:57 AM
ACCEPTED: X
REJECTED:
Operator: U02