
Midmark M11 - v1.0.5
Total Cycles: 4370

Name: RITTER2
Sterilizer ID: V2313228
02/21/2025 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 109.9 F 0.0
2:00 110.7 F 1.0
4:00 122.5 F 3.4
6:00 146.4 F 6.5
8:00 170.7 F 10.1
10:00 191.3 F 9.1
12:00 212.9 F 5.6
14:00 222.3 F 9.4
16:00 235.2 F 11.9
18:00 236.9 F 10.4
20:00 236.7 F 9.0
22:00 246.4 F 14.2
24:00 257.4 F 20.1
26:00 267.3 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.2
1:00 271.8 F 29.9
2:00 273.2 F 30.6
3:00 272.9 F 30.1
4:00 272.1 F 29.7

Min 271.0 F 29.0
Max 273.7 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.7 F 0.0
5:00 230.2 F 0.0
10:00 227.4 F 0.0
15:00 211.6 F 0.0
20:00 206.7 F 0.0
25:00 199.0 F 0.0
30:00 195.9 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:34
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:06:23
CYCLE COMPLETE

02/21/2025 02:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____