

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4111

Name: RITTER 1  
Sterilizer ID: V2178198  
12/13/2023 09:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.9 F 0.0  
2:00 75.9 F 0.2  
4:00 74.6 F 1.0  
6:00 93.4 F 3.0  
8:00 123.8 F 5.5  
10:00 150.2 F 8.4  
12:00 171.9 F 11.6  
14:00 206.8 F 2.6  
16:00 213.6 F 5.8  
18:00 226.8 F 9.1  
20:00 233.8 F 13.0  
22:00 233.3 F 8.8  
24:00 244.1 F 13.5  
26:00 241.4 F 11.3  
28:00 252.8 F 17.0  
30:00 262.9 F 22.9  
32:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.3 F 28.9  
1:00 272.6 F 29.1  
2:00 272.9 F 29.3  
3:00 272.4 F 29.4  
4:00 271.8 F 29.2  
  
Min 271.6 F 28.8  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
5:00 232.1 F 0.0  
10:00 220.6 F 0.0  
15:00 203.9 F 0.0  
20:00 202.6 F 0.0  
25:00 206.4 F 0.0  
30:00 210.5 F 0.0  
  
Drying Complete

FILLING: 1:00  
HEATING: 32:58  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:11:31  
CYCLE COMPLETE

12/13/2023 11:06 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_