

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1850

Name: RITTER2  
Sterilizer ID: V2313228  
12/15/2022 02:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 189.3 F 0.0  
2:00 189.7 F 3.5  
4:00 198.1 F 9.6  
6:00 217.4 F 6.5  
8:00 231.6 F 10.9  
10:00 234.0 F 9.1  
12:00 238.4 F 9.5  
14:00 249.7 F 15.8  
16:00 262.8 F 23.3  
18:00 271.9 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.7 F 29.5  
2:00 272.6 F 30.0  
3:00 272.2 F 29.3  
4:00 272.9 F 30.1  
  
Min 271.7 F 29.1  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.8 F 0.0  
5:00 204.3 F 0.0  
10:00 221.8 F 0.0  
15:00 230.1 F 0.0  
20:00 230.0 F 0.0  
25:00 227.1 F 0.0  
30:00 189.5 F 0.0  
  
Drying Complete

FILLING: 2:35  
HEATING: 18:04  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:57:51  
CYCLE COMPLETE

12/15/2022 03:46 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_