
Midmark M11 - v1.0.5
Total Cycles: 1064

Name: RITTER2
Sterilizer ID: V2313228
05/03/2022 12:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.3 F 0.0
2:00 150.1 F 2.1
4:00 167.9 F 6.0
6:00 193.9 F 10.9
8:00 220.2 F 5.0
10:00 233.8 F 10.2
12:00 233.8 F 8.2
14:00 243.2 F 12.7
16:00 241.7 F 11.3
18:00 254.3 F 17.8
20:00 265.6 F 24.6
22:00 271.5 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 271.7 F 29.2
2:00 271.6 F 29.2
3:00 272.1 F 29.3
4:00 272.9 F 29.5

Min 271.3 F 29.1
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 227.9 F 0.0
10:00 230.7 F 0.0
15:00 215.6 F 0.0
20:00 213.2 F 0.0
25:00 214.9 F 0.0
30:00 208.8 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:30
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:50
CYCLE COMPLETE

05/03/2022 01:34 PM
ACCEPTED: X
REJECTED:
Operator: U02