
Midmark M11 - v1.0.5
Total Cycles: 2333

Name: RITTER2
Sterilizer ID: V2313228
05/03/2023 02:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.3 F 0.0
2:00 146.8 F 2.3
4:00 164.7 F 5.9
6:00 189.3 F 10.7
8:00 211.1 F 5.0
10:00 224.6 F 10.2
12:00 229.1 F 7.8
14:00 240.5 F 12.8
16:00 240.7 F 11.6
18:00 253.6 F 18.2
20:00 264.8 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.5
1:00 272.5 F 29.9
2:00 271.7 F 29.8
3:00 272.4 F 29.9
4:00 271.5 F 29.7

Min 271.4 F 29.3
Max 272.7 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 198.9 F 0.0
10:00 215.0 F 0.0
15:00 221.7 F 0.0
20:00 225.9 F 0.0
25:00 228.5 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:53
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:30
CYCLE COMPLETE

05/03/2023 03:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____