
Midmark M11 - v1.0.5
Total Cycles: 512

Name: RITTER2
Sterilizer ID: V2313228
11/02/2021 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 91.5 F 0.0
2:00 95.8 F 0.0
4:00 112.7 F 0.1
6:00 146.3 F 1.0
8:00 161.9 F 4.2
10:00 183.7 F 7.9
12:00 199.9 F 11.8
14:00 217.0 F 5.7
16:00 227.5 F 9.7
18:00 235.2 F 10.9
20:00 239.1 F 11.4
22:00 236.9 F 9.5
24:00 249.7 F 15.6
26:00 260.4 F 21.7
28:00 270.0 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.5
2:00 271.5 F 29.5
3:00 272.1 F 29.4
4:00 271.5 F 29.4

Min 271.3 F 29.0
Max 272.5 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.1 F 0.0
5:00 225.4 F 0.0
10:00 221.4 F 0.0
15:00 208.6 F 0.0
20:00 200.5 F 0.0
25:00 202.4 F 0.0
30:00 199.0 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:55
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:07:06
CYCLE COMPLETE

11/02/2021 10:22 AM
ACCEPTED: X
REJECTED:
Operator: U01