

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1044

Name: RITTER 1  
Sterilizer ID: V2178198  
07/22/2021 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.8 F 0.0  
2:00 158.1 F 2.2  
4:00 177.3 F 5.6  
6:00 196.5 F 9.9  
8:00 223.7 F 5.7  
10:00 232.1 F 8.7  
12:00 241.7 F 12.1  
14:00 239.6 F 10.6  
16:00 236.2 F 8.6  
18:00 251.0 F 15.5  
20:00 262.9 F 22.4  
22:00 271.8 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.0  
1:00 272.4 F 28.8  
2:00 271.6 F 28.2  
3:00 271.2 F 28.1  
4:00 272.9 F 29.2

Min 271.0 F 27.9  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.6 F 0.0  
5:00 232.5 F 0.0  
10:00 225.6 F 0.0  
15:00 215.1 F 0.0  
20:00 202.0 F 0.0  
25:00 196.5 F 0.0  
30:00 193.7 F 0.0  
Drying Complete

FILLING: 1:33  
HEATING: 22:12  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:53  
CYCLE COMPLETE

07/22/2021 10:24 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02