
Midmark M11 - v1.0.0
Total Cycles: 1212

Name: RITTER 1
Sterilizer ID: V2178198
09/28/2021 11:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.8 F 0.0
2:00 179.2 F 3.3
4:00 204.9 F 7.6
6:00 225.5 F 13.5
8:00 232.7 F 9.1
10:00 238.0 F 9.5
12:00 244.8 F 12.5
14:00 242.3 F 11.2
16:00 255.7 F 18.1
18:00 267.6 F 25.3
20:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.6 F 28.7
2:00 272.6 F 29.1
3:00 271.9 F 28.9
4:00 272.1 F 28.7

Min 271.6 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.0 F 0.0
5:00 243.4 F 0.0
10:00 234.4 F 0.0
15:00 226.1 F 0.0
20:00 225.5 F 0.0
25:00 228.4 F 0.0
30:00 231.5 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 20:03
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:53
CYCLE COMPLETE

09/28/2021 12:11 PM
ACCEPTED: X
REJECTED:
Operator: U01