
Midmark M11 - v1.0.5
Total Cycles: 3176

Name: RITTER2
Sterilizer ID: V2313228
12/01/2023 12:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.3 F 0.0
2:00 137.9 F 0.8
4:00 153.3 F 3.6
6:00 178.2 F 7.3
8:00 196.4 F 11.5
10:00 214.4 F 5.2
12:00 225.3 F 9.7
14:00 233.6 F 9.9
16:00 237.7 F 11.5
18:00 236.7 F 9.8
20:00 249.7 F 15.9
22:00 260.8 F 22.2
24:00 270.9 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.8 F 30.2
2:00 272.3 F 29.7
3:00 271.2 F 29.2
4:00 271.8 F 30.0

Min 271.1 F 29.2
Max 273.2 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.6 F 0.0
5:00 218.8 F 0.0
10:00 217.9 F 0.0
15:00 222.3 F 0.0
20:00 221.7 F 0.0
25:00 224.2 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 24:48
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:22
CYCLE COMPLETE

12/01/2023 01:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____