
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001879
10:59 22/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:11
STERILIZING 3:12
136.4 C 226kPa 3:12
136.4 C 226kPa 3:42
136.4 C 226kPa 4:12
136.4 C 224kPa 4:42
136.4 C 227kPa 5:12
136.4 C 226kPa 5:42
136.4 C 224kPa 6:12
136.1 C 222kPa 6:42
136.1 C 222kPa 6:42
VENTING 6:42
AIR DRYING 7:37
CYCLE COMPLETE 37:37

Digital Signature #
E7ADC510AC778C05
