
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000739
9:31 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:32
PRESSURIZING 2:38
STERILIZING 3:40
136.4 C 232kPa 3:40
136.1 C 231kPa 4:10
136.4 C 233kPa 4:40
136.4 C 233kPa 5:10
136.3 C 232kPa 5:40
136.3 C 232kPa 6:10
136.3 C 233kPa 6:40
136.2 C 231kPa 7:10
136.2 C 231kPa 7:10
VENTING 7:10
AIR DRYING 8:03
CYCLE COMPLETE 8:18

Digital Signature #
B8816353D34F98FB
