
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000517
14:22 21/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:11
STERILIZING 4:22
136.0 C 230kPa 4:22
136.4 C 233kPa 4:52
136.4 C 233kPa 5:22
136.3 C 233kPa 5:52
136.1 C 230kPa 6:22
136.4 C 233kPa 6:52
136.3 C 232kPa 7:22
136.3 C 233kPa 7:52
136.3 C 233kPa 7:52
VENTING 7:52
AIR DRYING 8:46
CYCLE COMPLETE 23:46

Digital Signature #
B653235AA9F88699
