
Midmark M11 - v1.0.0
Total Cycles: 5345

Name: RITTER 1
Sterilizer ID: V2178198
02/19/2025 05:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.6 F 0.0
2:00 102.8 F 0.9
4:00 121.5 F 3.1
6:00 152.0 F 5.6
8:00 175.6 F 8.7
10:00 195.0 F 11.8
12:00 210.7 F 3.4
14:00 216.7 F 7.9
16:00 231.9 F 12.2
18:00 232.7 F 7.8
20:00 243.9 F 12.8
22:00 240.7 F 10.5
24:00 252.1 F 16.2
26:00 262.1 F 22.1
28:00 271.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.6
1:00 272.0 F 29.2
2:00 272.8 F 29.2
3:00 272.0 F 28.8
4:00 271.2 F 28.5

Min 271.0 F 28.4
Max 273.0 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.1 F 0.0
5:00 199.8 F 0.0
10:00 201.4 F 0.0
15:00 194.7 F 0.0
20:00 186.1 F 0.0
25:00 187.5 F 0.0
30:00 194.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:43
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:07:32
CYCLE COMPLETE

02/19/2025 07:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____