
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000183
10:08 17/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:34
PRESSURIZING 3:38
STERILIZING 5:11
136.0 C 220kPa 5:11
136.4 C 224kPa 5:41
136.4 C 224kPa 6:11
136.3 C 225kPa 6:41
136.2 C 223kPa 7:11
136.4 C 226kPa 7:41
136.3 C 224kPa 8:11
136.4 C 225kPa 8:41
136.4 C 225kPa 8:41
VENTING 8:41
AIR DRYING 9:44
CYCLE COMPLETE 24:44

Digital Signature #
0249B15598CD4E8F
