
Midmark M11 - v1.0.0
Total Cycles: 4313

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2024 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.0 F 0.0
2:00 70.5 F 0.2
4:00 78.4 F 1.1
6:00 111.2 F 3.2
8:00 140.6 F 5.9
10:00 168.7 F 9.1
12:00 187.0 F 12.7
14:00 211.4 F 3.3
16:00 216.6 F 6.4
18:00 225.5 F 9.8
20:00 236.0 F 11.9
22:00 234.8 F 9.2
24:00 242.4 F 11.6
26:00 243.0 F 11.6
28:00 254.3 F 17.4
30:00 264.2 F 23.2
32:00 271.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.9 F 28.8
2:00 272.6 F 29.2
3:00 272.0 F 28.4
4:00 272.8 F 29.1

Min 271.5 F 28.3
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 238.5 F 0.0
10:00 220.2 F 0.0
15:00 219.6 F 0.0
20:00 222.2 F 0.0
25:00 228.4 F 0.0
30:00 230.3 F 0.0

Drying Complete

FILLING: 0:53
HEATING: 32:55
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 01:11:17
CYCLE COMPLETE

02/21/2024 09:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____