

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3637

Name: RITTER2  
Sterilizer ID: V2313228  
05/23/2024 08:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.9 F 0.0  
2:00 70.0 F 0.0  
4:00 73.0 F 0.2  
6:00 95.9 F 1.6  
8:00 120.8 F 3.9  
10:00 153.1 F 6.7  
12:00 179.0 F 9.9  
14:00 225.8 F 13.3  
16:00 213.5 F 6.2  
18:00 222.8 F 9.7  
20:00 230.7 F 11.7  
22:00 235.0 F 9.5  
24:00 239.1 F 10.6  
26:00 243.4 F 12.4  
28:00 254.4 F 18.0  
30:00 264.1 F 24.0  
32:00 272.6 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.5  
1:00 272.8 F 29.9  
2:00 271.7 F 29.9  
3:00 271.5 F 29.4  
4:00 272.3 F 29.9  
  
Min 271.2 F 29.4  
Max 272.8 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.3 F 0.0  
5:00 219.7 F 0.0  
10:00 203.0 F 0.0  
15:00 174.5 F 0.0  
20:00 169.2 F 0.0  
25:00 165.0 F 0.0  
30:00 162.3 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:55  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:11:04  
CYCLE COMPLETE

05/23/2024 09:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_