

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2307

Name: RITTER 1  
Sterilizer ID: V2178198  
09/12/2022 07:36 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.7 F 0.0  
2:00 72.5 F 0.0  
4:00 80.8 F 0.0  
6:00 107.5 F 0.0  
8:00 142.0 F 0.0  
10:00 174.5 F 0.3  
12:00 196.3 F 0.7  
14:00 214.3 F 3.8  
16:00 229.9 F 6.6  
18:00 238.0 F 9.9  
20:00 239.0 F 11.6  
22:00 240.3 F 10.4  
24:00 236.3 F 8.3  
26:00 244.4 F 11.8  
28:00 246.0 F 13.0  
30:00 258.2 F 19.4  
32:00 268.8 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.9  
1:00 272.3 F 28.2  
2:00 272.6 F 28.7  
3:00 271.5 F 27.8  
4:00 272.4 F 28.4  
  
Min 271.5 F 27.8  
Max 272.7 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.3 F 0.0  
5:00 221.9 F 0.0  
10:00 228.1 F 0.0  
15:00 226.3 F 0.0  
20:00 246.6 F 0.0  
25:00 246.9 F 0.0  
30:00 230.6 F 0.0  
  
Drying Complete

FILLING: 0:44  
HEATING: 33:07  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:11:02  
CYCLE COMPLETE

09/12/2022 08:47 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_