

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4715

Name: RITTER 1  
Sterilizer ID: V2178198  
07/11/2024 09:52 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 148.3 F 0.0  
2:00 148.9 F 2.2  
4:00 153.9 F 5.2  
6:00 175.7 F 8.8  
8:00 194.8 F 12.7  
10:00 212.2 F 5.5  
12:00 221.6 F 9.9  
14:00 233.5 F 8.9  
16:00 237.0 F 10.7  
18:00 234.6 F 8.0  
20:00 247.6 F 14.4  
22:00 258.6 F 20.3  
24:00 268.4 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.4 F 29.4  
2:00 272.1 F 28.9  
3:00 271.1 F 28.5  
4:00 272.0 F 29.3

Min 271.1 F 28.5  
Max 272.5 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
5:00 191.7 F 0.0  
10:00 195.4 F 0.0  
15:00 197.0 F 0.0  
20:00 198.6 F 0.0  
25:00 208.0 F 0.0  
30:00 213.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 25:20  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:04:01  
CYCLE COMPLETE

07/11/2024 10:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_