
Midmark M11 - v1.0.0
Total Cycles: 1494

Name: RITTER 1
Sterilizer ID: V2178198
12/30/2021 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.2 F 0.0
2:00 201.1 F 3.1
4:00 218.1 F 8.3
6:00 236.6 F 9.1
8:00 237.2 F 11.4
10:00 238.1 F 9.4
12:00 236.7 F 8.5
14:00 250.7 F 15.3
16:00 263.2 F 22.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 27.9
1:00 272.4 F 28.7
2:00 272.2 F 28.4
3:00 272.8 F 28.8
4:00 272.1 F 28.9

Min 271.0 F 27.8
Max 273.0 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 218.8 F 0.0
10:00 194.1 F 0.0
15:00 191.6 F 0.0
20:00 189.2 F 0.0
25:00 187.7 F 0.0
30:00 190.7 F 0.0

Drying Complete

FILLING: 2:15
HEATING: 17:56
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:57:18
CYCLE COMPLETE

12/30/2021 12:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____