
Midmark M11 - v1.0.5
Total Cycles: 86

Name: RITTER2
Sterilizer ID: V2313228
06/10/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 74.5 F 0.2
4:00 78.2 F 1.2
6:00 96.0 F 3.1
8:00 128.4 F 5.6
10:00 161.2 F 8.6
12:00 176.6 F 12.2
14:00 199.2 F 4.2
16:00 217.8 F 7.7
18:00 224.4 F 11.6
20:00 228.9 F 7.5
22:00 239.5 F 12.5
24:00 238.9 F 10.6
26:00 249.9 F 16.2
28:00 260.3 F 21.9
30:00 269.3 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 272.9 F 30.0
1:00 272.2 F 30.3
2:00 271.9 F 29.9
3:00 272.8 F 30.3
4:00 272.4 F 30.0

Min 271.8 F 29.9
Max 273.1 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.1 F 0.0
5:00 214.9 F 0.0
10:00 207.5 F 0.0
15:00 204.6 F 0.0
20:00 201.5 F 0.0
25:00 201.4 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 31:34
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:09:54
CYCLE COMPLETE

06/10/2021 10:03 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____