

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000445  
8:34 15/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:27  
PRESSURIZING 4:28  
STERILIZING 5:51  
136.0 C 220kPa 5:51  
136.4 C 224kPa 6:21  
136.4 C 224kPa 6:51  
136.3 C 222kPa 7:21  
136.4 C 224kPa 7:51  
136.3 C 222kPa 8:21  
136.3 C 223kPa 8:51  
136.1 C 219kPa 9:21  
136.1 C 219kPa 9:21  
VENTING 9:21  
AIR DRYING 10:28  
CYCLE COMPLETE 10:50

-----  
Digital Signature #  
05E415BA3BC0A0E3  
-----