

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002157  
11:07 17/10/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:01  
PRESSURIZING 2:07  
STERILIZING 3:12  
136.0 C 220kPa 3:12  
136.4 C 223kPa 3:42  
136.4 C 223kPa 4:12  
136.4 C 225kPa 4:42  
136.3 C 224kPa 5:12  
136.4 C 223kPa 5:42  
136.1 C 223kPa 6:12  
136.4 C 225kPa 6:42  
136.4 C 225kPa 6:42  
VENTING 6:42  
AIR DRYING 7:32  
CYCLE COMPLETE 37:32  
-----

Digital Signature #  
31E42FA1DE8FEC73  
-----