
Midmark M11 - v1.0.0
Total Cycles: 5412

Name: RITTER 1
Sterilizer ID: V2178198
03/13/2025 12:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 120.8 F 0.0
2:00 123.4 F 1.5
4:00 134.5 F 4.3
6:00 157.3 F 7.6
8:00 179.4 F 11.1
10:00 212.9 F 4.7
12:00 212.8 F 6.5
14:00 224.2 F 10.5
16:00 233.8 F 7.8
18:00 238.5 F 10.9
20:00 235.3 F 8.2
22:00 247.6 F 14.4
24:00 258.3 F 20.2
26:00 268.2 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 271.4 F 28.9
2:00 272.2 F 29.6
3:00 272.7 F 29.4
4:00 272.4 F 29.2

Min 271.2 F 28.6
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 189.3 F 0.0
10:00 204.6 F 0.0
15:00 210.5 F 0.0
20:00 217.4 F 0.0
25:00 215.2 F 0.0
30:00 214.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 27:16
STERILIZING: 4:00
VENTING: 4:00
DRYING: 30:00
TOTAL CYCLE: 01:06:03
CYCLE COMPLETE

03/13/2025 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____