
Midmark M11 - v1.0.5
Total Cycles: 3705

Name: RITTER2
Sterilizer ID: V2313228
06/17/2024 10:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.8 F 0.0
2:00 175.3 F 1.8
4:00 177.2 F 5.3
6:00 200.1 F 10.2
8:00 213.6 F 6.5
10:00 226.1 F 8.0
12:00 237.0 F 12.5
14:00 236.9 F 10.0
16:00 237.9 F 9.2
18:00 248.2 F 14.7
20:00 259.3 F 20.9
22:00 269.3 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.9
1:00 272.6 F 29.5
2:00 271.5 F 28.9
3:00 272.6 F 29.7
4:00 271.8 F 29.0

Min 271.4 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 210.6 F 0.0
10:00 205.8 F 0.0
15:00 201.4 F 0.0
20:00 192.8 F 0.0
25:00 192.9 F 0.0
30:00 193.1 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 23:08
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:05
CYCLE COMPLETE

06/17/2024 11:09 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____