
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.9uS / 1.2ppm
CYCLE NUMBER 001871
11:34 18/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:03
PRESSURIZING 2:07
STERILIZING 3:13
136.0 C 224kPa 3:13
136.3 C 225kPa 3:43
136.2 C 224kPa 4:13
136.4 C 226kPa 4:43
136.4 C 226kPa 5:13
136.4 C 226kPa 5:43
136.4 C 225kPa 6:13
136.3 C 225kPa 6:43
136.3 C 225kPa 6:43
VENTING 6:43
AIR DRYING 7:32
CYCLE COMPLETE 37:32

Digital Signature #
4455C62E1E084265
