

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 006170  
7:54 28/08/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:57  
PRESSURIZING 3:01  
STERILIZING 4:05  
136.0 C 220kPa 4:05  
136.4 C 223kPa 4:35  
136.4 C 222kPa 5:05  
136.4 C 221kPa 5:35  
136.4 C 223kPa 6:05  
136.4 C 221kPa 6:35  
136.4 C 221kPa 7:05  
136.4 C 221kPa 7:35  
136.4 C 221kPa 7:35  
VENTING 7:35  
AIR DRYING 8:21  
CYCLE COMPLETE 13:00

-----  
Digital Signature #  
4CB35A6A3E89AB26  
-----