
Midmark M11 - v1.0.0
Total Cycles: 3885

Name: RITTER 1
Sterilizer ID: V2178198
10/10/2023 04:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.6 F 0.0
2:00 179.0 F 4.1
4:00 201.1 F 8.7
6:00 220.5 F 8.7
8:00 230.0 F 8.9
10:00 237.6 F 10.0
12:00 242.3 F 12.3
14:00 241.3 F 11.1
16:00 254.7 F 17.9
18:00 266.3 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 28.7
2:00 272.0 F 28.9
3:00 271.9 F 28.4
4:00 272.4 F 29.1

Min 271.2 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.3 F 0.0
5:00 257.5 F 0.0
10:00 247.5 F 0.0
15:00 225.1 F 0.0
20:00 223.4 F 0.0
25:00 232.5 F 0.0
30:00 226.2 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:59:08
CYCLE COMPLETE

10/10/2023 05:47 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____