

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.7uS / 1.0ppm  
CYCLE NUMBER 000808  
15:51 09/11/2021  
WRAPPED  
135 C for 10 min.  
CYCLE START 0:00  
CONDITIONING 1:48  
PRESSURIZING 3:42  
STERILIZING 4:38  
136.4 C 233kPa 4:38  
136.2 C 231kPa 5:08  
136.1 C 231kPa 5:38  
136.1 C 231kPa 6:08  
136.5 C 233kPa 6:38  
136.2 C 230kPa 7:08  
136.1 C 229kPa 7:38  
136.4 C 232kPa 8:08  
136.4 C 233kPa 8:38  
136.4 C 233kPa 9:08  
136.4 C 233kPa 9:38  
136.4 C 229kPa 10:08  
136.4 C 232kPa 10:38  
136.4 C 232kPa 11:08  
135.9 C 230kPa 11:38  
136.4 C 233kPa 12:08  
136.3 C 231kPa 12:38  
136.4 C 233kPa 13:08  
136.4 C 233kPa 13:38  
136.2 C 231kPa 14:08  
136.4 C 233kPa 14:38  
136.4 C 233kPa 14:38  
VENTING 14:38  
AIR DRYING 15:25  
CYCLE COMPLETE 30:25  
-----

Digital Signature #  
3298B5410149E86D  
-----