
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002016
7:49 12/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:44
PRESSURIZING 2:53
STERILIZING 3:59
136.0 C 221kPa 3:59
136.4 C 224kPa 4:29
136.4 C 224kPa 4:59
136.4 C 223kPa 5:29
136.4 C 224kPa 5:59
136.4 C 224kPa 6:29
136.3 C 224kPa 6:59
136.4 C 221kPa 7:29
136.4 C 221kPa 7:29
VENTING 7:29
AIR DRYING 8:21
CYCLE COMPLETE 33:46

Digital Signature #
E3E287EB31229DB0
