

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2637

Name: RITTER2  
Sterilizer ID: V2313228  
07/12/2023 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 190.6 F 0.0  
2:00 185.8 F 3.7  
4:00 201.3 F 8.6  
6:00 218.8 F 9.4  
8:00 230.1 F 10.1  
10:00 232.4 F 8.1  
12:00 241.3 F 11.5  
14:00 246.0 F 13.7  
16:00 258.7 F 20.6  
18:00 270.0 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.1  
1:00 272.2 F 29.2  
2:00 272.4 F 29.7  
3:00 271.6 F 28.9  
4:00 272.6 F 29.6  
  
Min 271.4 F 28.9  
Max 272.8 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.9 F 0.0  
5:00 212.5 F 0.0  
10:00 224.3 F 0.0  
15:00 224.3 F 0.0  
20:00 224.7 F 0.0  
25:00 224.0 F 0.0  
30:00 223.5 F 0.0  
  
Drying Complete

FILLING: 2:02  
HEATING: 18:52  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

07/12/2023 12:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_