
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006241
14:23 05/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:56
STERILIZING 2:49
136.0 C 220kPa 2:49
136.4 C 224kPa 3:19
136.4 C 220kPa 3:49
136.4 C 223kPa 4:19
136.1 C 220kPa 4:49
136.4 C 225kPa 5:19
136.2 C 221kPa 5:49
136.4 C 224kPa 6:19
136.4 C 224kPa 6:19
VENTING 6:19
AIR DRYING 7:10
CYCLE COMPLETE 7:13

Digital Signature #
1E0DEB30EE02C0E6
