
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 006510
13:27 20/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:42
STERILIZING 3:46
136.4 C 220kPa 3:46
136.2 C 222kPa 4:16
136.4 C 222kPa 4:46
136.4 C 225kPa 5:16
136.2 C 221kPa 5:46
136.4 C 223kPa 6:16
136.4 C 223kPa 6:46
136.2 C 224kPa 7:16
136.2 C 224kPa 7:16
VENTING 7:16
AIR DRYING 8:04
CYCLE COMPLETE 24:50

Digital Signature #
4C3758E31F575A92
