
Midmark M11 - v1.0.5
Total Cycles: 3873

Name: RITTER2
Sterilizer ID: V2313228
08/19/2024 02:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 94.5 F 0.0
2:00 96.1 F 0.0
4:00 106.1 F 0.3
6:00 143.8 F 2.2
8:00 173.8 F 4.9
10:00 190.0 F 8.2
12:00 201.8 F 11.7
14:00 217.5 F 4.1
16:00 226.0 F 7.7
18:00 235.9 F 11.7
20:00 233.7 F 8.5
22:00 245.2 F 13.5
24:00 243.4 F 12.3
26:00 254.3 F 17.9
28:00 264.3 F 23.9
30:00 272.4 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.2
1:00 272.8 F 29.5
2:00 272.0 F 29.7
3:00 271.6 F 29.1
4:00 272.7 F 29.4

Min 271.5 F 29.1
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 229.0 F 0.0
10:00 215.8 F 0.0
15:00 182.2 F 0.0
20:00 168.6 F 0.0
25:00 166.2 F 0.0
30:00 165.3 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:44
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:09:32
CYCLE COMPLETE

08/19/2024 03:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____