

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1635

Name: RITTER2  
Sterilizer ID: V2313228  
10/21/2022 09:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.4 F 0.0  
2:00 143.3 F 1.7  
4:00 176.8 F 4.8  
6:00 202.0 F 9.3  
8:00 215.7 F 6.8  
10:00 224.4 F 8.4  
12:00 236.3 F 13.5  
14:00 236.3 F 10.5  
16:00 236.7 F 8.5  
18:00 249.2 F 15.3  
20:00 260.4 F 22.0  
22:00 271.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.4  
1:00 272.0 F 29.6  
2:00 272.0 F 30.0  
3:00 271.7 F 29.3  
4:00 272.1 F 30.0

Min 271.3 F 29.3  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.6 F 0.0  
5:00 195.7 F 0.0  
10:00 213.7 F 0.0  
15:00 223.1 F 0.0  
20:00 227.4 F 0.0  
25:00 227.6 F 0.0  
30:00 225.1 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 22:39  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:00:49  
CYCLE COMPLETE

10/21/2022 10:00 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_