
Midmark M11 - v1.0.5
Total Cycles: 2602

Name: RITTER2
Sterilizer ID: V2313228
07/04/2023 12:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 209.9 F 0.0
2:00 207.4 F 4.1
4:00 208.4 F 9.8
6:00 222.8 F 6.9
8:00 234.7 F 11.3
10:00 235.1 F 9.1
12:00 238.6 F 9.7
14:00 248.7 F 15.0
16:00 260.8 F 21.8
18:00 271.9 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.5 F 29.4
2:00 271.5 F 29.0
3:00 272.6 F 29.5
4:00 271.6 F 28.8

Min 271.5 F 28.8
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.8 F 0.0
5:00 220.5 F 0.0
10:00 234.7 F 0.0
15:00 230.4 F 0.0
20:00 225.6 F 0.0
25:00 223.7 F 0.0
30:00 222.0 F 0.0

Drying Complete

FILLING: 2:28
HEATING: 18:34
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

07/04/2023 01:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____