
Midmark M11 - v1.0.0
Total Cycles: 4685

Name: RITTER 1
Sterilizer ID: V2178198
06/27/2024 01:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.9 F 0.0
2:00 100.6 F 0.8
4:00 109.4 F 2.8
6:00 136.3 F 5.5
8:00 162.3 F 8.8
10:00 184.3 F 12.6
12:00 207.3 F 4.8
14:00 221.4 F 8.7
16:00 232.4 F 12.8
18:00 232.3 F 8.9
20:00 240.8 F 11.3
22:00 242.3 F 12.0
24:00 253.9 F 17.8
26:00 264.3 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.5 F 29.6
2:00 272.5 F 29.6
3:00 272.3 F 29.7
4:00 272.0 F 29.6

Min 270.9 F 28.7
Max 273.1 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 218.8 F 0.0
10:00 212.0 F 0.0
15:00 218.7 F 0.0
20:00 222.5 F 0.0
25:00 226.8 F 0.0
30:00 228.6 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 27:60
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:06:39
CYCLE COMPLETE

06/27/2024 02:44 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____