
Midmark M11 - v1.0.5
Total Cycles: 4064

Name: RITTER2
Sterilizer ID: V2313228
10/29/2024 03:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.4 F 0.0
2:00 96.0 F 0.0
4:00 115.7 F 0.0
6:00 156.8 F 0.1
8:00 195.8 F 1.8
10:00 196.4 F 5.3
12:00 207.9 F 8.5
14:00 218.8 F 12.0
16:00 227.5 F 6.1
18:00 237.3 F 10.5
20:00 234.8 F 8.6
22:00 244.8 F 13.1
24:00 242.9 F 11.8
26:00 254.5 F 17.7
28:00 264.7 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 271.9 F 29.4
2:00 272.6 F 29.2
3:00 272.1 F 29.6
4:00 272.7 F 29.3

Min 270.9 F 28.4
Max 273.1 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.4 F 0.0
5:00 232.0 F 0.0
10:00 182.5 F 0.0
15:00 180.5 F 0.0
20:00 176.7 F 0.0
25:00 174.7 F 0.0
30:00 173.2 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 29:59
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

10/29/2024 04:23 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____