
Midmark M11 - v1.0.0
Total Cycles: 2074

Name: RITTER 1
Sterilizer ID: V2178198
06/29/2022 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.1 F 0.0
2:00 133.4 F 0.4
4:00 153.0 F 2.5
6:00 179.1 F 5.2
8:00 207.7 F 9.1
10:00 218.4 F 10.4
12:00 227.6 F 7.9
14:00 239.1 F 12.6
16:00 235.5 F 9.3
18:00 241.2 F 9.7
20:00 246.4 F 13.4
22:00 258.2 F 19.8
24:00 269.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 29.1
2:00 272.0 F 28.4
3:00 271.7 F 28.7
4:00 272.6 F 29.0

Min 271.4 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.2 F 0.0
5:00 211.4 F 0.0
10:00 231.9 F 0.0
15:00 236.6 F 0.0
20:00 228.9 F 0.0
25:00 226.7 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 25:11
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:03:11
CYCLE COMPLETE

06/29/2022 01:17 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____